

Annexure (August 2021)

Post publication of CGSI Electronic pulse oximeter testing report in (May 2021), ***Theo Solutions Pvt Ltd. (Serial No 11, in the above said report)***, have intimated us that they collected samples of their oximeters from across India and a batch of ten samples were sent to Godrej & Boyce Manufacturing Co. Ltd., Mumbai for a calibration testing. According to recent test results, submitted to CGSI, the SpO₂ % and Prbpm now are in line with the specified standards.

CGSI truly appreciates the company's quick actions and efforts to investigate and test oximeter batch samples and take steps to resolve any technical issue of their products, within a limited span of time.